

## TB-2715 TERMINAL BLOCK

This installation guide describes how to install and connect sign is to the TB-2715 terminal block for use with PXI-653X and PXI-660X n. idules.

Use the TB-2715 terminal block to connect to the signals of a PXI-653X or PXI-660X module. Connect your signals to the screw terminals housed in the terminal block then install the TB-2715 to the PXI front panel.

## What You Need to Get Started

To set up and use the TB-2715, you will need the following:	То
☐ TB-2715 terminal block	
Quick-reference labels for PXI-653X and PXI-660X	
<ul><li>☐ One of the following:</li><li>– PXI-653X module</li><li>– PXI-660X module</li></ul>	
☐ 653X User Manual or 660X User Manual	
□ 0.10 in. slotted screwdriver	
☐ Number 1 Phillips screwdriver	
☐ Wire cutters	
☐ Wire insulation stripper	

## **Connecting Signals to Your Terminal Block**

Refer to Figure 1 as you perform the following steps to connect signals from the peripheral device to the TB-2715. For more information about connecting signals to your terminal block, see your 653X User Manual or 660X User Manual.

National Instruments™, ni.com™, and PXI™ are trademarks of National Instruments Corporation. Product and company names mentioned herein are trademarks or trade names of their respective companies.